


*ANNEX XII***CE MARKING OF CONFORMITY**

The CE conformity marking shall consist of the initials CE taking the following form~



- If the marking is reduced or enlarged the proportions given in the above graduated drawing must be respected.
- The various components of the CE marking must have substantially the same vertical dimension, which may not be less than **5** mm.

This minimum dimension may be waived for small scale devices.